



CERTIFICATE OF ANALYSIS

Nova Na/pH Conditioning Solution
Lot Number 25051023

Ref Number 00018
Exp Date 2026-05-20
Date of Manufacture 2025-02-20
UDI# (01)00385480000185(11)250220(17)260520(10)25051023

	<u>Spec</u>	<u>Pass/Fail</u>
pH	≤ 5.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Awishi M

Approval

QC Inspector

Title

2/27/2025

Date