



CERTIFICATE OF ANALYSIS

Nova 1/4/5/8/10/11/12/13/14 Reference Sensor
Lot Number 25053033

Part Number 14244
Exp Date 2026-08-22

Date of Manufacture 2025-02-22

UDI# (01)00385480142441(11)250222(17)260822(10)25053033

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	≥ 9.8	Pass
Set A Level 3	137.2 - 143.2	Pass
CI Slope	≥ 10.5	Pass
Set A Level 3	110.4 - 118.4	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

²Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

2/28/2025

Date