



CERTIFICATE OF ANALYSIS

Prime ES Plus Cal Cartridge 300 Sample
Lot Number 25058050

Ref Number 62227
Exp Date 2026-06-23
Date of Manufacture 2024-12-23

UDI# (01)00385480622271(11)241223(17)260623(10)25058050

Cal A	L/N	24358083		Cal B	L/N	24358084		Reference	L/N	24358086	
		<u>Spec</u>	<u>Pass/Fail</u>			<u>Spec</u>	<u>Pass/Fail</u>			<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L		134.1 - 135.9	Pass	Na mmol/L		76.5 - 77.5	Pass	KCl		1.8 - 2.2	Pass
K mmol/L		4.02 - 4.08	Pass	K mmol/L		9.80 - 10.20	Pass				
Cl mmol/L		114.1 - 115.9	Pass	Cl mmol/L		44.6 - 45.5	Pass				
iCa mmol/L		0.99 - 1.01	Pass	iCa mmol/L		1.98 - 2.02	Pass				
iMg mmol/L		0.39 - 0.49	Pass	iMg mmol/L		1.48 - 1.52	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Approval

QC Inspector
Title

3/6/2025
Date