

*nova*  
*biomedical*  
**CERTIFICATE OF ANALYSIS**


Prime Reference Cartridge  
Lot Number 25060040

Part Number 42043  
Exp Date 2026-09-01  
Date of Manufacture 2025-03-01  
UDI# (01)00385480420433(11)250301(17)260901(10)25060040

|           | <u>Spec</u>  | <u>Pass/Fail</u> |
|-----------|--------------|------------------|
| Na+ Slope | 9.50 - 12.00 | Pass             |
| K+ Slope  | 9.00 - 11.50 | Pass             |
| Hct Slope | 2.00 - 3.00  | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer



Approval

QC Inspector  
Title

3/6/2025  
Date