



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 400 Sample w. Creat/BUN
Lot Number 25063038

Ref Number 57834
Exp Date 2026-08-19

Date of Manufacture 2025-02-19

UDI# (01)00385480578349(11)250219(17)260819(10)25063038

Table with 3 columns: Cal A L/N 25052038, Cal B L/N 25050014, Cal C L/N 25050015. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, pH, tHb g/dL, and Assigned Value\*/Result.

Table with 3 columns: Cal D L/N 25052041, Flush L/N 25050038, Reference L/N 25052037. Rows include pH @ 37C, PO2, and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

3/12/2025
Date