



## CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pCO2 Sensor  
Lot Number 25049063

Ref Number 48600  
Exp Date 2027-02-18

Date of Manufacture 2025-02-18

UDI# (01)00385480486002(11)250218(17)270218(10)25049063

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.40 - 11.60	Pass
Std C mV	0.00 ± 150.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector

Title

3/14/2025

Date