

OI-922

Page 29 of 67

Rev. BM

Vialed and Packaged Test Strip Release


Test Strips



## CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

|  |  |
|--|--|
| Customer:  | NIA  |
| Description:   | STATSENSOR-I CREAT Test Strips                                   |
| Catalog Number:  | P/N 43272  |
| Nova Biomedical P/N  | P/N 43272  |
| Control Level 1: CV (%) $\leq$ 15%<br>Control Level 2: CV (%) $\leq$ 10%<br>Control Level 3: CV (%) $\leq$ 10% | PASS / FAIL<br>PASS / FAIL<br>PASS / FAIL                        |
| Qty:   | 3,780  |
| Lot No.  | 4925059129   |
| Mfg. Date:   | 2025-02-28   |
| Expiration Date:   | 2026-02-28   |
| Unique Device Identifier (UDI):  | (01)20385480432720<br>(11)250228<br>(17)260228<br>(10)4925059129 |

  
Quality Control Manager

3/10/25  
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454