

OI-922

Page 27 of 67

Rev. BM

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS/FAIL PASS/FAIL PASS/FAIL
Qty:	① 20,016 22,014
Lot No.	4925 055129
Mfg. Date:	2025-02-24
Expiration Date:	2026-02-24
Unique Device Identifier (UDI):	(01)20385480433703 (11)250224 (17)260224 (10)4925055129

Man Browne
Quality Control Manager

3/10/25
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454

① MFC 3/10/25