

## **CERTIFICATE OF ANALYSIS**

Allegro HbA1c Test Cartridge Lot Number 25066017

tridge Ref Number 54215 Exp Date 2026-03-07 Date of Manufacture 2025-03-07 UDI# (01)00385480542159(11)250307(17)260307(10)25066017

	Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail	
HbA1c %	4.9 - 6.7	Pass	HbA1c %	8.1 - 11.7	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Muyht

Approval

<u>QC Inspector</u> Title

<u>3/14/2025</u> Date