



**CERTIFICATE OF ANALYSIS**

STP Prime Plus Reference Cartridge  
Lot Number 25065059

Ref Number 57823

Exp Date 2026-09-06

Date of Manufacture 2025-03-06

UDI# (01)00385480578233(11)250306(17)260906(10)25065059

| Internal QC Level 4 | <u>Spec</u>   | <u>Pass/Fail</u> |
|---------------------|---------------|------------------|
| Na, mmol/L          | 139.3 - 147.3 | Pass             |
| K, mmol/L           | 3.73 - 4.23   | Pass             |
| CL, mmol/L          | 123.4 - 132.4 | Pass             |
| External QC Level 4 | <u>Spec</u>   | <u>Pass/Fail</u> |
| Na, mmol/L          | 134.3 - 142.3 | Pass             |
| K, mmol/L           | 3.73 - 4.23   | Pass             |
| CL, mmol/L          | 118.9 - 127.9 | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

3/19/2025

Date