



CERTIFICATE OF ANALYSIS

STP Prime Plus VET Cal-Cartridge 500 Sample
Lot Number 25063037

Ref Number 58396
Exp Date 2026-08-19
Date of Manufacture 2025-02-19

Table with 3 columns: Cal A L/N 25052038, Cal B L/N 25050014, Cal C L/N 25050015. Rows include HCO3, Na, K, Ca, Mg, Glu, Lac, BUN, pH, and tHb with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25052041, Flush L/N 25050038, Reference L/N 25052037. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

3/13/2025

Date