



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Plus Glucose Sensor
Lot Number 25063025

Part Number 27809
Exp Date 2027-03-04

Date of Manufacture 2025-03-04

UDI# (01)00385480278096(11)250304(17)270304(10)25063025

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	76 - 90	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Houghton", written in a cursive style.

Approval

QC Supervisor
Title

3/21/2025
Date