



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Cal Cartridge 1 w/Creat  
Lot Number 25066053

Ref Number 48836  
Exp Date 2026-08-13  
Date of Manufacture 2025-02-13

UDI# (01)00385480488365(11)250213(17)260813(10)25066053

Cal A L/N 25052032

	<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	7.351 - 7.357	Pass
Lac mmol/L	13.5 - 16.5	Pass

Cal B L/N 25044017

	<u>Spec</u>	<u>Pass/Fail</u>
pH @ 37C	6.822 - 6.828	Pass
Na mmol/L	69.5 - 70.5	Pass
K mmol/L	9.9 - 10.1	Pass
Cl mmol/L	39.7 - 40.3	Pass
Ca mmol/L	1.97 - 2.03	Pass
Mg mmol/L	1.48 - 1.52	Pass

Cal C L/N 25044018

	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass
K mmol/L	3.97 - 4.03	Pass
Cl mmol/L	118.0 - 120.0	Pass
BUN mg/dl	9.9 - 10.1	Pass
Glu mg/dl	78.8 - 81.2	Pass
Lac mmol/L	1.8 - 2.2	Pass
Ca mmol/L	0.99 - 1.02	Pass
Mg mmol/L	0.49 - 0.51	Pass
Creat mg/dl	0.98 - 1.02	Pass
HCO3 mmol/L	18.34 - 19.08	Pass

Cal D L/N 25058014

	<u>Spec</u>	<u>Pass/Fail</u>
BUN mg/dl	49.4 - 50.6	Pass
Glu mg/dl	197 - 203	Pass
Lac mmol/L	5.5 - 6.5	Pass
Creat mg/dl	9.8 - 10.2	Pass
CO3 mmol/L	36.68 - 38.18	Pass

Flush L/N 25057024

	<u>Spec</u>	<u>Result</u>
Aqueous Salt Solution	N/A	Present

Reference L/N 25051008

	<u>Spec</u>	<u>Pass/Fail</u>
KCl mol/L	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*

Approval

QC Inspector

Title

3/20/2025

Date