



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 25070010

Part Number 48554
Exp Date 2026-06-11
Date of Manufacture 2025-03-11
UDI# (01)00385480485548(11)250311(17)260611(10)25070010

| | <u>Spec*</u> | <u>Mean Result</u> |
|------------------------|--------------|--------------------|
| Slope | ≥ 15.00 | 26.64 |
| pHox Ultra/CCX Ctl L4 | 1.7 - 2.3 | 1.9 |
| pHox Ultra/CCX Ctl L5 | 6.3 - 7.7 | 7.2 |
| Lin Soln Set A Level 1 | 14.0 - 20.0 | 16.7 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a faint horizontal line.

Approval

QC Inspector
Title

3/21/2025
Date