



CERTIFICATE OF ANALYSIS

BioProfile Control Level 9

Lot Number 25055004

Catalog Number 43473

Exp Date 2026-05

	Spec	Pass/Fail
Particles/mL	$55.00 \times 10^5 - 65.00 \times 10^5$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

3/26/2025

Date