



CERTIFICATE OF ANALYSIS

Prime ES Comp Plus Cal Cartridge 500 Sample
Lot Number 25058051

Ref Number 62229
Exp Date 2026-06-23

Date of Manufacture 2024-12-23

UDI# (01)00385480622295(11)241223(17)260623(10)25058051

Cal A L/N 24358085

	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.1 - 135.9	Pass
K mmol/L	4.02 - 4.08	Pass
Cl mmol/L	114.1 - 115.9	Pass
iCa mmol/L	0.99 - 1.01	Pass
iMg mmol/L	0.37 - 0.47	Pass

Cal B L/N 24358084

	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	76.5 - 77.5	Pass
K mmol/L	9.80 - 10.20	Pass
Cl mmol/L	44.6 - 45.5	Pass
iCa mmol/L	1.98 - 2.02	Pass
iMg mmol/L	1.48 - 1.52	Pass

Reference L/N 24358086

	<u>Spec</u>	<u>Pass/Fail</u>
KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

3/25/2025

Date