



## CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Lactate Sensor  
Lot Number 25072028

Part Number 48553  
Exp Date 2027-03-13

Date of Manufacture 2025-03-13

UDI# (01)00385480485531(11)250313(17)270313(10)25072028

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHox Ctl L3	76 - 90	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M.", written in a cursive style.

Approval

QC Inspector  
Title

3/26/2025  
Date