



CERTIFICATE OF ANALYSIS

Stat Profile pHox Sodium Sensor

Lot Number 25072040

Ref Number 21523

Exp Date 2026-03-13

Date of Manufacture 2025-03-13

UDI# (01)00385480215237(11)250313(17)260313(10)25072040

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 - 11.00	Pass
Std A mV	10.00 - 130.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a faint horizontal line.

Approval

QC Inspector

Title

3/26/2025

Date