

OI-922

Page 65 of 67

Rev. BM

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSTRIP CAPILLARY LACTATE 50 COUNT, PACKAGED
Catalog Number:	61440
Nova Biomedical P/N	61440
<u>Target Range</u>	<u>CV (%)</u> <u>SD (mmol/L)</u>
0.8-2.5 mmol/L	≤ 6.5% ±0.3
3.0-5.0 mmol/L	≤ 5.5% ±0.5
10-14 mmol/L	≤ 5.0% ±1.2
	PASS / FAIL
	PASS / FAIL
	PASS / FAIL
Qty:	299
Lot No.	2624274249
Mfg. Date:	2024-09-30
Expiration Date:	2026-09-30
Unique Device Identifier (UDI):	(01)20385480614409 (11) 240930 (17) 260930 (10) 2624274249

Mauborne
Quality Control Manager

3/6/25
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454

Released Confidential