



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample
Lot Number 25065053

Ref Number 53464
Exp Date 2026-08-24
Date of Manufacture 2025-02-24
UDI# (01)00385480534642(11)250224(17)260824(10)25065053

Cal A	L/N	25055026		
			<u>Assigned Value*</u>	<u>Result</u>
pH			7.354	N/A
pCO2 mmHg			24.7	N/A

Cal B	L/N	25055027		
			<u>Spec</u>	<u>Pass/Fail</u>
pH			6.837 - 6.843	Pass
pO2 mmHg			167.0	N/A

Flush	L/N	25055057		
			<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg			50.7	N/A
pO2 mmHg			101.1	N/A

Reference	L/N	25055054		
			<u>Spec</u>	<u>Pass/Fail</u>
KCl			1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

03/20/2025
Date