



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 300 Sample w. Creat/BUN
Lot Number 25069072

Ref Number 57833
Exp Date 2026-08-12

Date of Manufacture 2025-02-12

UDI# (01)00385480578332(11)250212(17)260812(10)25069072

Table with 3 columns: Cal A L/N 25057016, Cal B L/N 25057017, Cal C L/N 25057018. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, and tHb g/dL with Spec, Pass/Fail, Assigned Value*, and Result columns.

Table with 3 columns: Cal D L/N 25057019, Flush L/N 25043022, Reference L/N 25057020. Rows include pH @ 37C, PO2, and KCl with Spec, Pass/Fail, Assigned Value*, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

3/18/2025

Date