



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.															
Description:	TEST STRIPS, NIPRO STATSTRIP™ KETONE, PACKAGED 20 CT															
Catalog Number:	54288															
Nova Biomedical P/N	54288															
<table><thead><tr><th>Target Range</th><th>MPE</th><th>SD/CV (%)</th></tr></thead><tbody><tr><td>0.1-0.8 mmol/L</td><td>+/- 0.3 mmol/L</td><td>≤ 0.1 mmol/L</td></tr><tr><td>1.5-2.5 mmol/L</td><td>+/- 20%</td><td>≤ 10%</td></tr><tr><td>3.5-5.0 mmol/L</td><td>+/- 20%</td><td>≤ 10%</td></tr></tbody></table>	Target Range	MPE	SD/CV (%)	0.1-0.8 mmol/L	+/- 0.3 mmol/L	≤ 0.1 mmol/L	1.5-2.5 mmol/L	+/- 20%	≤ 10%	3.5-5.0 mmol/L	+/- 20%	≤ 10%	<table><tbody><tr><td><u>PASS</u> / FAIL</td></tr><tr><td><u>PASS</u> / FAIL</td></tr><tr><td><u>PASS</u> / FAIL</td></tr></tbody></table>	<u>PASS</u> / FAIL	<u>PASS</u> / FAIL	<u>PASS</u> / FAIL
Target Range	MPE	SD/CV (%)														
0.1-0.8 mmol/L	+/- 0.3 mmol/L	≤ 0.1 mmol/L														
1.5-2.5 mmol/L	+/- 20%	≤ 10%														
3.5-5.0 mmol/L	+/- 20%	≤ 10%														
<u>PASS</u> / FAIL																
<u>PASS</u> / FAIL																
<u>PASS</u> / FAIL																
Qty:	893															
Lot No.	2225073249															
Mfg. Date:	2025-03-14															
Expiration Date:	2027-03-14															

Mac Broome
Quality Control Manager

3/25/2025
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454