

nova
biomedical
CERTIFICATE OF ANALYSIS


STP® pHox Ultra™ / CCX™ pH Sensor
Lot Number 25072007

Part Number 48580
Exp Date 2026-09-13
Date of Manufacture 2025-03-13
UDI# (01)00385480485807(11)250313(17)260913(10)25072007

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer



Approval

QC Inspector
Title

4/2/2025
Date