



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane  
Lot Number 25078007

Part Number 48552

Exp Date 2026-06-19

Date of Manufacture 2025-03-19

UDI# (01)00385480485524(11)250319(17)260619(10)25078007

|                       | <u>Spec*</u>  | <u>Mean Result</u> |
|-----------------------|---------------|--------------------|
| Slope                 | 11.00 - 45.00 | 37.68              |
| pHox Ultra/CCX Ctl L4 | 73 - 87       | 79                 |
| pHox Ultra/CCX Ctl L5 | 255 - 305     | 291                |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. Ross", with a stylized flourish at the end.

Approval

QC Inspector

Title

3/27/2025

Date