



CERTIFICATE OF ANALYSIS

Nova 1/4/5/8/10/11/12/13/14 Reference Sensor
Lot Number 25084038

Part Number 14244
Exp Date 2026-09-25

Date of Manufacture 2025-03-25

UDI# (01)00385480142441(11)250325(17)260925(10)25084038

	<u>Spec*</u>	<u>Pass/Fail</u>
Na Slope	≥ 9.8	Pass
Set A Level 3	137.2 - 143.2	Pass
CI Slope	≥ 10.5	Pass
Set A Level 3	110.4 - 118.4	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

²Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

4/3/2025

Date