



CERTIFICATE OF ANALYSIS

STP pHox Plus Chloride Sensor
Lot Number 25084016

Part Number 27812
Exp Date 2026-06-25
Date of Manufacture 2025-03-25
UDI# (01)10385480278123(11)250325(17)260625(10)25084016

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.00	Pass
Std A mV	55.00 - 95.00	Pass
pHox Ctl L1	120.0 - 129.0	Pass
pHox Ctl L3	79.6 - 89.6	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "John", written over a faint, light-colored circular stamp.

Approval

QC Inspector
Title

4/4/2025
Date