



CERTIFICATE OF ANALYSIS

Linearity Std Set A Level 6
Lot Number 25076014

Ref Number 14071
Exp Date 2026-09-17
Date of Manufacture 2025-03-17
UDI# (01)00385480140713(11)250317(17)260917(10)25076014

| | <u>Spec</u> | <u>Pass/Fail</u> |
|------------|---------------|------------------|
| Na mmol/L | 98.0 - 102.0 | Pass |
| K+ mmol/L | 98.0 - 102.0 | Pass |
| Cl- mmol/L | 196.0 - 204.0 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Awishi M

Approval

QC Inspector

Title

3/28/2025

Date