

OI-922

Page 29 of 67

Rev. BM

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) \leq 15% Control Level 2: CV (%) \leq 10% Control Level 3: CV (%) \leq 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	9,270
Lot No.	4924337129
Mfg. Date:	2024 -12 -02
Expiration Date:	2025 -12 -02
Unique Device Identifier (UDI):	(01)20385480432720 (11)241202 (17)251202 (10)4924337129

Marc Brune
Quality Control Manager

12/18/24
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454