



## CERTIFICATE OF ANALYSIS

Reagent Pack Autosampler Flex 2  
Lot Number 25065036

Part Number 62450  
Exp Date 2026-02-10  
Date of Manufacturing 2025-02-10

Solution DP L/N	25041067	Solution S L/N	25041066
NaOCl %	<u>Spec</u> 0.43 - 0.73	pH	<u>Spec</u> 6.900 - 7.100
	<u>Pass/Fail</u> Pass		<u>Pass/Fail</u> Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO<sub>3</sub>, pH, KCl, iCa, NH<sub>4</sub>, K.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*

Approval

QC Inspector

Title

3/28/2025

Date