



## CERTIFICATE OF ANALYSIS

STP pHox Reference Sensor  
Lot Number 25083066

Ref Number 21520  
Exp Date 2026-09-24

Date of Manufacture 2025-03-24

UDI# (01)00385480215206(11)250324(17)260924(10)25083066

	<u>Spec</u>	<u>Pass/Fail</u>
Na Slope	9.00 - 11.00	Pass
ADT 4	>= 220	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in dark ink, appearing to read "P. Hernandez", written in a cursive style.

Approval

QC Inspector  
Title

4/7/2025  
Date