



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 100 Sample
Lot Number 25083111

Ref Number 57825
Exp Date 2026-09-04

Date of Manufacture 2025-03-04

UDI# (01)00385480578257(11)250304(17)260904(10)25083111

Table with 3 columns: Cal A L/N 25071034, Cal B L/N 25071035, Cal C L/N 25071037. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN and tHb g/dL with Spec, Pass/Fail, and Result columns.

Table with 3 columns: Cal D L/N 25071036, Flush L/N 25063010, Reference L/N 25070037. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval: Awishi M
QC Inspector
Title
Date: 4/7/2025