



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 25087103

Ref Number 53359
Exp Date 2026-09-17
Date of Manufacture 2025-03-17
UDI# (01)00385480533591(11)250317(17)260917(10)25087103

Cal A	L/N	25080007	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.353	N/A
	pCO2 mmHg	25.1	N/A

Cal B	L/N	25080008	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	25076015	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	51.1	N/A
	pO2 mmHg	100.6	N/A

Reference	L/N	25080009	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

04/08/2025
Date