



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 25092013

Part Number 48554
Exp Date 2026-07-02
Date of Manufacture 2025-04-02
UDI# (01)00385480485548(11)250402(17)260702(10)25092013

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	45.72
pHox Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.0
Lin Soln Set A Level 1	14.0 - 20.0	18.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

4/7/2025
Date