

*nova*  
*biomedical*  
**CERTIFICATE OF ANALYSIS**


Prime Reference Cartridge  
Lot Number 25095008

Part Number 42043  
Exp Date 2026-10-05  
Date of Manufacture 2025-04-05  
UDI# (01)00385480420433(11)250405(17)261005(10)25095008

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ Slope	9.50 - 12.00	Pass
K+ Slope	9.00 - 11.50	Pass
Hct Slope	2.00 - 3.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer



Approval

QC Inspector  
Title

4/9/2025  
Date