



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ Potassium Sensor
Lot Number 25092019

Ref Number 48582

Exp Date 2026-07-02

Date of Manufacture 2025-04-02

UDI# (01)00385480485821(11)250402(17)260702(10)25092019

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.00 to 11.00	Pass
Std C mV	0 ± 50.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a faint horizontal line.

Approval

QC Inspector

Title

4/10/2025

Date