



## CERTIFICATE OF ANALYSIS

Bio pHox Calibrator Cartridge  
Lot Number 25087052

Ref Number 41874  
Exp Date 2026-09

Cal A L/N 25069065			Cal B L/N 25069017			Cal C L/N 25065043		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HCO <sub>3</sub> mmol/L	139.0 - 141.0	Pass	HCO <sub>3</sub> mmol/L	39.2 - 40.8	Pass	pH @ 37C	7.352 - 7.358	Pass
Na mmol/L	19.6 - 20.4	Pass						
Cal D L/N 25065042			Reference L/N 25065027					
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>			
pH @ 37C	6.802 - 6.808	Pass	KCl mol/L	1.8 - 2.2	Pass			
Na mmol/L	69.5 - 70.5	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

04/14/2025  
Date