



CERTIFICATE OF ANALYSIS

Stat Profile pHox/Basic Calibrator Cartridge A
Lot Number 25087053

Ref Number 23935
Exp Date 2026-09-06
Date of Manufacture 2025-03-06
UDI# (01)00385480239356(11)250306(17)260906(10)25087053

| | | | | | | | | |
|-------------------------|---------------|------------------|-------------------------|-------------|------------------|--------------------|---------------|------------------|
| Cal A L/N 25069065 | | | Cal B L/N 25069017 | | | Cal C L/N 25065043 | | |
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| HCO ₃ mmol/L | 20.1 - 20.9 | Pass | HCO ₃ mmol/L | 40.2 - 41.8 | Pass | pH @ 37C | 7.342 - 7.348 | Pass |
| Na mmol/L | 139.3 - 141.3 | Pass | | | | | | |
| Cal D L/N 25065042 | | | Reference L/N 25065027 | | | | | |
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> | | | |
| pH @ 37C | 6.792 - 6.798 | Pass | KCl mol/L | 1.8 - 2.2 | Pass | | | |
| Na mmol/L | 71.5 - 72.5 | Pass | | | | | | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

04/11/2025
Date