

CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge Lot Number 25097008

Ref Number 54652 Exp Date 2026-10-07 Date of Manufacture 2025-04-07 UDI# (01)00385480546522(11)250407(17)261007(10)25097008

Ctl Level 1

Ctl Level 2

Spe	ec <u>Pass/</u>	Fail	Spec	Pass/Fail
TCHOL mg/dL 146	- 53 Pas - 216 Pas - 124 Pas	s TCHOL mg/dL	294 - 410	Pass Pass Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

fuman Achta

Approval

QC Inspector Title

<u>4/12/2025</u> Date