



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane  
Lot Number 25087013

Part Number 48552  
Exp Date 2026-06-28  
Date of Manufacture 2025-03-28  
UDI# (01)00385480485524(11)250328(17)260628(10)25087013

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	30.85
pHox Ultra/CCX Ctl L4	73 - 87	79
pHox Ultra/CCX Ctl L5	255 - 305	286

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. S. S. S.", written over a horizontal line.

Approval

QC Inspector  
Title

4/15/2025  
Date