



## CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pH Sensor  
Lot Number 25091065

Ref Number 21522  
Exp Date 2027-04-01  
Date of Manufacture 2025-04-01  
UDI# (01)00385480215220(11)250401(17)270401(10)25091065

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	10.0 - 10.8	Pass
Std A mV	0.00 +/- 110.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "John", written over a horizontal line.

Approval

QC Inspector  
Title

4/15/2025  
Date