



CERTIFICATE OF ANALYSIS

STP pHox Plus Reference Sensor
Lot Number 25098020

Ref Number 28954
Exp Date 2026-10-08
Date of Manufacture 2025-04-08
UDI# (01)00385480289542(11)250408(17)261008(10)25098020

	<u>Spec</u>	<u>Pass/Fail</u>
Na Slope	9.00 - 11.00	Pass
ADT 4	≥ 220.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "P. Hernandez", written over a light gray rectangular background.

Approval

QC Inspector
Title

4/17/2025
Date