



CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor
Lot Number 25105021

Part Number 48591
Exp Date 2028-04-15
Date of manufacture 2025-04-15
UDI#(01)00385480485913(11)250415(17)280415(10)25105021

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	8.77	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. M. ...", written over a faint horizontal line.

Approval

QC Inspector
Title

4/18/2025
Date