



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pCO2 Sensor
Lot Number 25092030

Ref Number 21524

Exp Date 2027-04-02

Date of Manufacture 2025-04-02

UDI (01)00385480215244(11)250402(17)270402(10)25092030

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|-----------------|------------------|
| Slope | 9.50 - 11.30 | Pass |
| Std A mV | 0.00 +/- 150.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in dark ink, appearing to be "John", written in a cursive style.

Approval

QC Inspector

Title

4/21/2025

Date