

nova
biomedical

CERTIFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	47940
Nova Biomedical P/N	47940
Target Range (mg/dL) SD (mg/dL) /CV (%)	
30-50 ≤ 4 mg/dL	<u>PASS</u> / FAIL
51-110 ≤ 6.0%	<u>PASS</u> / FAIL
111-150 ≤ 5.0%	<u>PASS</u> / FAIL
151-250 ≤ 5.0%	<u>PASS</u> / FAIL
251-400 ≤ 5.0%	<u>PASS</u> / FAIL
Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method ≥ 95%	<u>PASS</u> / FAIL
Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method ≥ 98%	<u>PASS</u> / FAIL
Qty:	304
Lot No.	2125098309
Mfg. Date:	2025.04.08
Expiration Date:	2027.10.08

Marc Browne
Quality Control Manager

4/15/25
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454

OI-922

Page 57 of 67

Rev. BN

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2

RELEASE TEST RESULTS

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2125098309

(Spiked whole blood) Target Range:30-50		
Mean (mg/dL) N=100	SD (mg/dL) ≤ 4	YSI (reference, mg/dL)
35	1.9	38.7
(Spiked whole blood) Target Range:51-110		
Mean (mg/dL) N=100	CV% ≤ 6.0	YSI (reference, mg/dL)
92	5.3	97.3
(Spiked whole blood) Target Range:111-150		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
135	4.1	144.0
(Spiked whole blood) Target Range:151-250		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
215	2.4	229.0
(Spiked whole blood) Target Range:251-400		
Mean (mg/dL) N=100	CV% ≤ 5.0	YSI (reference, mg/dL)
307	2.5	382.5

Accuracy Criteria 1: Total within +/-12 mg/dL or +/- 12% of reference method $\geq 95\%$	Accuracy Criteria 2: Total within +/-15 mg/dL or +/- 15% of reference method $\geq 98\%$
97.0%	99.6%

Marc Broome
Quality Control Manager
Nova Biomedical Corporation

4/15/25
Date
200 Prospect Street Waltham, MA 02454

Released Confidential