

OI-922

Page 66 of 67

Rev. BN

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA																		
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED																		
Catalog Number:	47486																		
Nova Biomedical P/N	47486																		
<table border="1"><thead><tr><th>Target Range</th><th>CV (%)</th><th>SD (mmol/L)</th></tr></thead><tbody><tr><td>0.8-2.5 mmol/L</td><td>≤ 6.5%</td><td>±0.3</td></tr><tr><td>3.0-5.0 mmol/L</td><td>≤ 5.5%</td><td>±0.5</td></tr><tr><td>10-14 mmol/L</td><td>≤ 5.0%</td><td>±1.2</td></tr></tbody></table>	Target Range	CV (%)	SD (mmol/L)	0.8-2.5 mmol/L	≤ 6.5%	±0.3	3.0-5.0 mmol/L	≤ 5.5%	±0.5	10-14 mmol/L	≤ 5.0%	±1.2	<table border="1"><tbody><tr><td>PASS</td><td>FAIL</td></tr><tr><td>PASS</td><td>FAIL</td></tr><tr><td>PASS</td><td>FAIL</td></tr></tbody></table>	PASS	FAIL	PASS	FAIL	PASS	FAIL
Target Range	CV (%)	SD (mmol/L)																	
0.8-2.5 mmol/L	≤ 6.5%	±0.3																	
3.0-5.0 mmol/L	≤ 5.5%	±0.5																	
10-14 mmol/L	≤ 5.0%	±1.2																	
PASS	FAIL																		
PASS	FAIL																		
PASS	FAIL																		
Qty:	3,516																		
Lot No.	2625090249																		
Mfg. Date:	2025-03-31																		
Expiration Date:	2027-03-31																		
Unique Device Identifier (UDI):	(01)20385480 474867 (11) 250331 (17) 270331 (10) 2625090249																		


Quality Control Manager

4/15/25
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454