

## **CERTIFICATE OF ANALYSIS**

Prime ES plus Cal Cartridge 300 Sample Lot Number 25098046

Ref Number 62227 Exp Date 2026-09-21

Date of Manufacture 2025-03-21

UDI# (01)00385480622271(11)250321(17)260921(10)25098046

Cal A L/N 25080012		Cal E	Cal B L/N 25080010				Reference L/N 25080035		
<u>Spec</u>	Pass/Fail		<u>Sp</u>	<u>ec</u>	Pass/Fail		<u>Spec</u>	Pass/Fail	
Na mmol/L 134.1 - 135.9	Pass	Na mmol/L	76.5 -	77.5	Pass	KCI mol/L	1.8 - 2.2	Pass	
K mmol/L 4.02 - 4.08	Pass	K mmol/L	9.80 -	10.20	Pass				
CI mmol/L 114.1 - 115.9	Pass	CI mmol/L	44.6 -	45.5	Pass				
iCa mmol/L 0.99 - 1.01	Pass	iCa mmol/L	1.98 -	2.02	Pass				
iMg mmol/L 0.38 - 0.48	Pass	iMg mmol/L	1.48 -	1.52	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Haneen Aldarwish

Title

04/22/2025

Date