



## CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0425072301

Exp Date 2027-09-13

Date of Manufacture 2025-03-13

UDI # (01)00385480417419(11)250313(17)270913(10)0425072301

	Spec	Mean Result
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*A. M.*

Approval

QC Inspector

Title

4/16/2025

Date