



CERTIFICATE OF ANALYSIS

NCRT 4 Calibrator Pack
Lot Number 509941

Ref Number 15647
Exp Date 2026-06-27
Date of Manufacture 2025-03-27
UDI#(01)00385480156479(11)250327(17)260627(10)509941

CAI A L/N 25091024			Cal B L/N 25091023		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	129.1-130.9	Pass	Na mmol/L	59.4-60.6	Pass
K mmol/L	3.97-4.03	Pass	K mmol/L	9.8-10.2	Pass
Cl mmol/L	115.1-116.9	Pass	Cl mmol/L	60.4-61.6	Pass
HCO3 mmol/L	19.2-20.8	Pass	HCO3 mmol/L	9.6-10.4	Pass
Li mmol/L	0.49-0.51	Pass	Li mmol/L	4.95-5.05	Pass
Cal H L/N 25092017			Reference L/N 25086008		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
CO2 Releasing Agent	N/A	Present	KCl mol/L	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Haneen Aldarwish

Approval

QC Inspector
Title

4/18/2025
Date