



## CERTIFICATE OF ANALYSIS

Nova 5 Calibrator Pack  
Lot Number 501328

Ref Number 15648  
Exp Date 2027-01-13  
Date of Manufacture 2025-01-13  
UDI# (01)00385480156486(11)250113(17)270113(10)501328

Cal A	L/N	25013028	Cal B	L/N	25013028	Reference	L/N	25013028
	<u>Spec</u>	<u>Pass/fail</u>		<u>Spec</u>	<u>Pass/fail</u>		<u>Spec</u>	<u>Pass/fail</u>
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass	KCl mol/L	1.8 - 2.2	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass			
Cl mmol/L	113.6 - 115.4	Pass	Cl mmol/L	45.5 - 46.5	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

4/22/2025  
Date