



CERTIFICATE OF ANALYSIS

Glucose pHOX plus Membrane Kit
Lot Number 25106009

Part Number 28969
Exp Date 2026-07-16

Date of Manufacture 2025-04-16

UDI# (01)00385480289696(11)250416(17)260716(10)25106009

	<u>Spec*</u>	<u>Mean Result</u>
Slope	5.00 - 45.00	24.41
pHOX plus L3	78 - 92	86
Set A L1	≥ 470	512

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

04/25/2025

Date